

STRYKER Wheelchair

TopGun
Mobility
WHEELS FOR LIFE



USER MANUAL

Table of Contents

INTRODUCTION	1
INTENDED USE	1
SCHEMATIC AND FEATURES	2
SPECIFICATION TABLE.....	3
OPERATIONAL INSTRUCTIONS - USE.....	4
SAFETY PRECAUTIONS	9
MOUNTING	11
CARE and MAINTENANCE	14
STORAGE	15
PERIOD OF USE.....	15
WARRANTY	16

INTRODUCTION

CONGRATULATIONS on the purchase of your new Top Gun Wheelchair. We are excited to provide you one of our innovative wheelchairs.

Consult your doctor or health professional before using a mobility aid. He/She will advise on the most appropriate use for you. This User Manual contains important information concerning the proper maintenance and safe operation of your wheelchair. Please read this manual thoroughly and carefully prior to operating your new wheelchair to familiarise yourself with all operational instructions.

INTENDED USE

The Stryker Wheelchair is intended for indoor or outdoor use with or without an attendant. This foldable wheelchair is designed for carrying a **maximum weight capacity up to 135kg** with comfort with easy-to-use features.

MANDATORY! Do not operate your new Stryker Wheelchair for the first time without reading and understanding this user manual completely. Healthcare Professionals and/or the authorized dealer to give instructions for the use of the product.


NOTE:

Check all parts for shipping damage. In case of shipping damage, Do NOT use.
Contact authorised Top Gun Mobility dealer for further instruction.

PLEASE NOTE :

All details contained within this user manual are correct as at time of print. Specifications/Recommendations are subject to change without prior notice. This manual covers the use of various models, you will note that many of the operations and adjustments are similar and as such images and some information may depict a different model to that which you have purchased.

SCHEMATIC AND FEATURES



- Painted aluminium frame
- Swing away footrest design to reduce accidents during user transfer.
- Front castor and rear wheel height adjustable
- Quick Release on rear wheels
- Safe Belt
- Double Cross Bar
- Height adjustable, forward & backward armrest with PU pad
- Half folding backrest design makes for easy transportation and compact storage.
- PU Tyres
- Anti-tip Wheels
- Tension Seat and Back
- Attendant hand brakes ensure safety

SPECIFICATION TABLE



Backrest Height & Inclination	45.5cm
Folded Width	34.5cm
Gross Weight	17kg
Height (Seat to floor)	46cm
Overall Dimensions (W x D x H)	67cm x 99cm x 90cm
Seat Depth	42cm
Seat Width	45cm
Weight Capacity	135kg
Wheel Measurements – Front (Diameter)	20cm (8")
Wheel Measurements – Rear (Diameter)	61cm (24")


OPERATIONAL INSTRUCTIONS - USE

For the safe use of your wheelchair, in and out of the house, you need to be confident in the main manoeuvres, such as pushing, braking, going back and forth on flat surfaces (can be done with the help of an assistant). Avoid slopes, uneven surfaces etc.

Wheelchair use


Make sure that the front wheels are in the front and the rear wheels braked when sitting on the wheelchair especially when you are wanting to transfer to another location (e.g. chair, bed, etc.).

Apply the brake to the wheels every time you raise from the wheelchair, and if necessary, ask for help from an assistant or family members.



How to get out of the wheelchair alone

Avoid bending too far down, or sideways, or backward, as there would be a risk of overturning or losing balance especially if the chair has a reduced seat and / or a high seat height (e.g. cushion). Make sure that the front wheels are in the front and the rear wheels braked when sitting on the wheelchair transferring to another location (e.g. chair, bed, etc.). Keep in balance!!!




Overcoming Obstacles

Ask for an assistant, especially if you have to overcome obstacles larger than 5 cm. Unassisted: the wheelchair can tip over and causing damage to the person and the chair; the chair may become stuck or wheels may be suspended in the air.




Upward Slope

Do not hang anything above the backrest or above the armrest. This may cause the chair to tip over. To go up hills, install the anti-tippers. To this end insert the anti-tippers into the rear tube and take care that the safety clip protrudes out of the hole in the back tube, locking the anti-tippers into place.



Downward Slope

If you are not confident, please ask for help when you are going down a slope. Be careful that the wheels don't become stuck at the bottom of the slope when going down a slope.



Stairs – Up and Down

When stairs are the only option, once the chair is tilted backwards by a person holding the chair by the handles, rest the rear wheels on the steps. The second person must hold a fixed part of the front area of the frame, lift the seat and hold it until the first person stands on the next step and so on. Do not release the chair until you have reached the last step and moved away from the steps. Crossing narrow steps: When manoeuvring through narrow spaces, such as building entrances, doors, etc... Watch out for elbows. - Danger of injury!



Using an Elevator/Lift

Automatic door closing can cause hand injuries !

Transport of the Wheelchair

Although the method to be followed for transport of the Chair is not unified.

1. Remove the footrests.
2. Remove the armrests.
3. Remove the wheels using the quick release.
4. Fold the chair by pulling the handles on the sides of the seat until the chair is completely folded.
5. Put it in a car.

How To Propel The Chair Yourself

Once the user is seated in the chair, take the rings of the rear wheels with both hands, and forcefully push forward so that the chair begins to move. Release the rings. When the chair stops, repeat the action again.



Using the Brakes

Once the chair is stationary, the user holds the brakes with their hands, activates them, so that the rear wheels are completely locked.



SAFETY PRECAUTIONS



Failure to follow the safety precautions may result in the risk of serious injury.

DO NOT exceed the maximum supported weight of 135kg.

- Make sure that all parts are in good working order.
- Do not use the product on stairs or escalators.
- Only use the accessories and parts supplied by the manufacturer.
- Check the correct operation of the brakes before use.
- This wheelchair has been designed for transport inside vehicles of all kinds. Always transfer the user from the wheelchair to a vehicle seat that has been approved by the vehicle manufacturer and use all available safety restraints for both the user and the wheelchair.
- Do not use this wheelchair on roads, streets, or shared areas with other surface traffic.
- Be sure to fully activate the wheel brakes when getting out of the wheelchair, or when you are in a lift or wheelchair lift. The wheelchair must be on a flat, stable, and level surface.
- When getting in or out of the wheelchair, do not climb or put any weight on your footrest. When you try to extend your arms over your back, do so only until your arm can be extended without changing your sitting position.
- Do not stand in the wheelchair
- If you must lean or lean on the chair, make sure that the front wheels are pointing forward, so that the base of the wheel extends to stabilise the chair.
- Seek professional advice when going up or down slopes of more than 10 cm. If you must stop on a slope, avoid abrupt and rapid movements when moving. Movement can cause instability due to the angle of the chair. The use of brakes can lead to accidental braking, uncontrolled movement, or a sudden stop. Try not to change direction on a ramp, as this may cause instability and cause the chair to tip over.
- During use, lock the brakes for all actions.
- All wheels must always be in contact with the ground.

- When the chair is in a stationary position, the brakes must be locked. Do not attempt to move the chair if the brakes are locked.
- Make sure that the footrests are properly engaged.
- Make sure the removal pin is securely engaged in the rear wheels and frame of the chair.
- Be careful when climbing pavements or walking/operating on sloping, uneven or slippery surfaces. Failure to do so may result in a significant risk of falling or injury.
- Do not use the product on a wet, icy or snowy slope.
- If any parts are damaged, broken, worn, or missing, contact the dealer.

Stability - The Stryker Wheelchair is designed to ensure safety and stability, provided that you maintain the center of gravity while using it. To preserve the chair's stability, it's important to avoid sudden movements that could lead to a shift in your center of gravity. To ensure the safe operation of the wheelchair, these instructions must be followed.

Hand Brakes - For safety, this wheelchair comes with handbrakes to provide better control.

Inclined Surface - When attempting to navigate inclined surfaces, including ramps, with this wheelchair, it's important to consider your strength, capabilities, and limitations. For correct techniques, a qualified healthcare professional may be able to provide product demonstrations and valuable advice.

NEVER put your hands, fingers, feet or legs into any moving part. **When making adjustments** (i.e. adjusting the leg rest height etc.), always take care that your hands or fingers are not trapped between any of the moving parts



LOAD RESTRICTIONS – Maximum carrying capacity is 135kg.



TYRE INFLATION – Only if your wheelchair is equipped with pneumatic tyres. If your Stryker Wheelchair is equipped with pneumatic tyres, check the tyre pressure at least once a week. Proper inflation pressures will prolong the life of your tyres and help ensure smooth operation.

It is advisable to use the chair on flat surfaces, inside or outside a room, on longer journeys it is preferable that there is always a companion to help the user, in case of encountering obstacles.

In order to use the wheelchair without problems, the speed must be between 3 and 5 km/h. The brakes are used to stop the chair. **Do not use the brakes if you are moving**, especially if you go up or down a slope, there is a risk of tipping over.

Users should not stand on the footrests as they could tip the chair over and fall.

While using the wheelchair on an incline, do not tip the chair or turn suddenly.

Do not brake if you are going downhill, you could tip over.

MOUNTING

Seat height

The seat height is adjusted with the rear wheel support bracket where the axle is located. Once the seat height is adjusted at the rear, adjust the front part correctly so that the seat is at the correct angle (not tilted too far forward or too far back).

Change the fork or change the position of the front wheel by choosing another hole in the fork. The purpose of adjusting the seat height is to maximise the user's efficiency by adapting the seat to the environment according to the user's physical and mechanical needs.

Adjusting footrests – installation, adjustment and removal

The Footrests are folding inward for easy transfer. The footrest is adjustable in degrees of inclination. Note: To remove the footrest, squeeze the trigger toward the chair, rotate the footrests outward or inward and raise them. To install the footrests, place the plastic swivel on the tubing in front of the seat, turn the footrests inward, until you grasp the locking bolt.



Fixing system-adjusting rear wheel

The fixing-adjustment system is the most important adjustment for the wheelchair because it determines the center of gravity, the rotation of the chair, the degrees of the wheels and the height of the seat. The center of gravity is determined by the user's position on the chair. After changing the height of the seat, you may need to adjust the height of the attendant brake handle.

NOTE: the higher the seat, the more unstable the wheelchair can be.

The position of the rear axle with respect to the horizontal coordinate of the center of gravity determines the ease of propulsion and handling of the chair, but also the subsequent stability.

Armrests

To allow easy access from each side of the chair, the armrests are folded down. Check that they are locked when they are replaced.

Options and accessories

- Small transit wheel
- Quick release
- Seat cushion

CARE and MAINTENANCE



Failure to care for and monitor your product may result in the risk of serious injury.

- Clean the product regularly with warm water and a soft cloth. Do not use abrasive products, detergents or solvents.
- Dry with a clean, dry soft cloth.
- Make sure there is no accumulation of dirt or debris around the axle, brakes or wheels.
- Do not use cleaning products containing phenol or chlorine, as this could damage aluminium and polyamide materials.
- Lubricate the wheels with a little oil if they squeak.
- Periodically check the wear and strength of all screws and bolts. Replace all broken, damaged or worn parts immediately.
- Check that the brakes are working before each use. Check the brakes by pushing the handles into the locked position, then trying to push the wheelchair.
- Check the upholstery has no cuts, is not torn or heavily used. Replace if necessary.
- If there is a malfunction, do not repair the product yourself. Consult an authorized dealer.

STORAGE



Failure to observe the storage conditions may lead to deterioration of the product and thus to the risk of serious injury.

Do not store your product for an extended period of time near a heat source or in the sun (e.g. behind a window or near a radiator) or near a cold source.

- Keep away from all flames and sources of sparks.
- Observe the conditions for storing and stowing your product :
 - In a dry and temperate place
 - Package your product to protect it from dust, corrosion (e.g. abrasive elements, sand, sea water, salty air).
 - Store all removed parts together in the same place to avoid mixing them with other products when reassembling.
 - All components must be stored without any load applied (do not place any objects that are too heavy on the product elements resulting in bending or warping of the components).

PERIOD OF USE

The life expectancy of this product is 5 years under normal use, safety and maintenance conditions. Beyond this period, the product can be used for as long as it is still in good condition.

Modifications

We have designed and engineered your Stryker Wheelchair to provide maximum mobility and utility. Under no circumstances should you modify, add, remove, or disable any feature, part, or function of your wheelchair.



WARNING! Do not modify your Stryker Wheelchair in any way not authorized by the manufacturer. Do not use accessories if they have not been tested or approved by the manufacturer.

WARRANTY

This is to certify that your TOP GUN MOBILITY product offers a 5 Year Limited Warranty from the date of original purchase. Should a defect or fault arise within this period the Dealer, from where the product was originally purchased, should be notified immediately. Failure to register the fault may invalidate the warranty.

The Dealer from where the product was purchased is responsible for any warranty repairs. Top Gun Mobility warrants that if a fault occurs within this period due to faulty workmanship or materials that the product will be repaired (or replaced at the Company's discretion) free of charge.

Any unauthorised modifications will invalidate the warranty.

FIVE-YEAR LIMITED WARRANTY Structural frame ONLY	12 MONTH LIMITED WARRANTY Your Top Gun Stryker Wheelchair is fully guaranteed for 12-months from the date of purchase against faults arising due to defects in manufacture or materials. This warranty does not detract from but is in addition to your legal rights.
--	---

COMMERCIAL USE

Any Top Gun Mobility products used in a commercial application will only be covered by a 12-month warranty from the purchase date and will be excluded from any coverage mentioned in the warranty relating to 5 (five) year warranty. All other conditions in this warranty will apply.

NOT COVERED UNDER WARRANTY

The guarantee does not extend to those items which may need replacement due to normal wear and tear or damage to the product caused by misuse or accident for which Top Gun Mobility or its agent cannot be held responsible. This warranty does not include service calls.



Copyright Top Gun Mobility
Owner's Manual & Warranty Booklet 2023